



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/550,297	09/21/2005	Raymond Andrieu	LEMAN.001APC	1931
20995 7590 10/30/2008 KNOBBE MARTENS OLSON & BEAR LLP 2040 MAIN STREET FOURTEENTH FLOOR IRVINE, CA 92614				
EXAMINER SCHILLINGER, ANN M				
ART UNIT		PAPER NUMBER		
3774				
NOTIFICATION DATE		DELIVERY MODE		
10/30/2008		ELECTRONIC		

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

jcartee@kmob.com  
eOAPilot@kmob.com

# Office Action Summary

**Application No.**

10/550,297

**Applicant(s)**

ANDRIEU ET AL

**Examiner**

ANN SCHILLINGER

**Art Unit**

3774

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 16 July 2008.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 21-35 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 21-35 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SF/ICE)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

### **DETAILED ACTION**

#### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(c) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 21-23, 28, and 31-35 are rejected under 35 U.S.C. 102(e) as being anticipated by Moe et al. (US Pub. No. 2003/0023302). Moe et al. discloses the following of the claimed invention: a biological prosthesis comprising an aortic valve obtained from an animal (11; paragraphs 0004-0005); a tubular wall (12); and a reinforcement device (16) comprising a rod (13) implanted in the tubular outer wall. Claims 22, 23, 28, and 31-35 are disclosed as shown in Figures 1A and 1B.

#### ***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 24-26 are rejected under 35 U.S.C. 103(a) as being unpatentable over Moe et al. in view of Rosen (US Pat. No. 4,345,340). Moe et al. teaches the invention substantially as claimed, however, Moe et al. does not teach the use of commissure points. Rosen teaches a heart

valve with commissure points in col. 3, lines 5-15 for the purpose of securing the tubular wall to the reinforcement device. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the device of Moe et al. by using commissure points to secure the tubular wall to the reinforcement device.

Claim 27 is rejected under 35 U.S.C. 103(a) as being unpatentable over Moe et al. in view of Gupta (US Pat. No. 4,851,000). Moe et al. teaches the invention substantially as claimed, however, Moe et al. does not teach coating the reinforcement device with a fluoropolymer material. Gupta teaches a heart valve with a fluoropolymer material coating on the reinforcement device in col. 4, lines 40-50 for the purpose of making the prosthesis more durable. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the device of Moe et al. by using a fluoropolymer material coating on the reinforcement device in order to make the prosthesis more durable.

Claims 29 and 30 are rejected under 35 U.S.C. 103(a) as being unpatentable over Moe et al. in view of Pavcnik et al. (US Pub. No. 2001/0039450). Moe et al. teaches the invention substantially as claimed, however, Moe et al. does not teach the intraparietal reinforcement devices having helical shapes. Pavcnik et al. teaches an implantable valve with an intraparietal reinforcement device that has a helical shape in Figures 2-8 and paragraphs 0008-0010 for the purpose of helping the prosthesis to better conform to the rounded contours of the vessel it is placed in. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the device of Moe et al. by giving the intraparietal reinforcement devices helical shapes in order to help the prosthesis to better conform to the rounded contours of the vessel it is placed in.

***Response to Arguments***

In view of the amendments submitted on 7/16/2008, the claim objection of claim 27 is withdrawn.

Applicant's arguments with respect to claims 21-35 have been considered but are moot in view of the new ground(s) of rejection.

***Conclusion***

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to ANN SCHILLINGER whose telephone number is (571)272-6652. The examiner can normally be reached on Mon. thru Fri. 9 a.m. to 4 p.m..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, David Isabella can be reached on (571) 272-4749. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/A. S./  
Examiner, Art Unit 3774  
/William H. Matthews/  
Primary Examiner, Art Unit 3774